

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

DIESSE

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME 88000 TOTAL IgE

Code	88000		
Lot No.	952	 2024-04	 2023-07

REAGENTS

Label	Description	Lot.	Exp.	NO
D5771	88000- TOTAL IgE	952	2024-04	6x6
B5771	CALIBRATORE 88000- TOTAL IgE	952	2024-09	1
C5771	CONTROL SERUM 88000- TOTAL IgE	952	2024-09	1

CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	350	≥ 100
Control Serum (KIU/l)	200,0	130,0-302,6

CALIBRATOR

9	1	1	9	5	2	5	2	3	0	4	9	3	9	5	3	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature 




CONTROL SERUM

9	1	1	9	5	2	5	1	8	3	4	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature 



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents 

DATE 20.10.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.