

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME LKM-1 CHORUS

Code 86080

Lot No. 970-A



2024-09



2023-10

REAGENTS

Label	Description	Lot.	Exp.	NO
D3641	LKM-1	970	2024-09	6X6
B3641	CALIBRATORE LKM-1	970-1	2024-10	1
C3641	CONTROL SERUM LKM-1	970-1	2024-10	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	590	≥ 100
Control Serum AU/mL	40,8	26,7-67,9

CALIBRATOR

7	0	3	9	7	0	6	4	1	9	3	3	5	5	5	9	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature HB



CONTROL SERUM


7	0	3	9	7	0	6	1	3	2	4	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature HB



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... HB

DATE 20/10/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive