

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME Jo-1 - CHORUS

Code	86030		
Lot No.	978	2025-01	2023-10

REAGENTS

Label	Description	Lot.	Exp.	NO
D3261	Jo-1	978	2025-01	6X6
B3261	CALIBRATORE Jo-1	978	2025-01	1
C3261	CONTROL SERUM Jo-1	978	2025-01	1

CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	390	≥ 100
Control Serum (AU/mL)	62,5	35,0-92,0

CALIBRATOR

1	9	7	9	7	8	7	1	1	9	3	1	3	4	4	1	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature MB



CONTROL SERUM

1	9	7	9	7	8	7	2	1	3	9	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature MB



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents. Bura

DATE 20/10/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive.