

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME GLIADIN-G - CHORUS

Code 86056

Lot No. 972



2024-09



2023-09

REAGENTS

Label	Description	Lot.	Exp.	NO
D3411	Gliadin-G	972	2024-09	6X6
B3411	CALIBRATORE Gliadin-G	972	2024-09	1
C3411	CONTROL SERUM Gliadin-G	972	2024-09	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

CALIBRATOR (mOD)	560	≥ 100
CONTROL SERUM (AU/ml)	56,3	28,1-111,8

CALIBRATOR


5	0	2	9	7	2	6	3	1	8	2	3	5	4	4	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *MB*



CONTROL SERUM


5	0	2	9	7	2	6	1	3	2	4	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *MB*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *MB*

DATE 20/10/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive