

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME RESPIRATORY SYNCYTIAL VIRUS IgG - CHORUS

Code 81036

Lot No. 970



2024-09



2023-09

REAGENTS

Label	Description	Lot.	Exp.	NO
D7661	Respiratory Syncytial Virus IgG	970	2024-09	6X6
B7661	CALIBRATORE Respiratory Syncytial Virus IgG	970	2024-09	1
C7661	CONTROL SERUM Respiratory Syncytial Virus IgG	970	2024-09	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	506	≥ 100
Control Serum (Ndx)	1,6	1,2-2,2
Negative Control (Ndx)	0,0	$\leq 0,7$

CALIBRATOR

1	5	9	9	7	0	6	0	0	0	0	0	6	0	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader

YES



Operator Signature

88



CONTROL SERUM

1	5	9	9	7	0	6	1	2	2	2	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader

YES



Operator Signature

88



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE

25.10.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent,