

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

DIESSE

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME EPSTEIN-BARR EARLY ANTIGEN IgG

Code 81058

Lot No. 979



2025-01



2023-10

REAGENTS

Label	Description	Lot	Exp.	NO
D2372	Epstein-Barr Early Antigen IgG	979	2025-01	6x6
B2373	CALIBRATORE Epstein-Barr Early Antigen IgG	979	2025-01	1
C2372	CONTROL SERUM Epstein-Barr Early Antigen IgG	979	2025-01	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	310	≥ 100
Control Serum (Index)	4,2	2,8-5,2
Negative Control (Index)	0,0	$< 0,7$

CALIBRATOR


0	0	7	9	7	9	7	0	0	0	0	0	2	5	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *HB*

CONTROL SERUM

0	0	7	9	7	9	7	2	8	5	2	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *HB*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *[Signature]*

DATE 08/11/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.