

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME Epstein-Barr EBNA IgG -CHORUS

Code 81057

Lot No. 971



2024-09



2023-09

REAGENTS

Label	Description	Lot.	Exp.	NO
D2361	Epstein-Barr EBNA IgG	971	2024-09	6X6
B2362	CALIBRATORE Epstein-Barr EBNA IgG	971	2024-09	1
C2361	CONTROL SERUM Epstein-Barr EBNA IgG	971	2024-09	1

CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	300	≥ 100
Control Serum (INDEX)	3,3	1,8-3,4
Negative Control (INDEX)	0,0	$< 0,7$

CALIBRATOR

0 0 6 9 7 1 6 0 0 0 0 0 3 9 0 0 0 0


Label checked with barcode reader YES ☒ Operator Signature *HB*



CONTROL SERUM

0 0 6 9 7 1 6 1 8 3 4 0 0 0 0 0 0 1

Label checked with barcode reader YES ☒ Operator Signature *HB*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *HB*

DATE 08.11.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.