

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

DIESSE

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME TOXOCARA IgG -CHORUS

Code 81280

Lot No. 978

2024-10

2023-10

REAGENTS

Label	Description	Lot.	Exp.	NO
D0020	TOXOCARA IgG	978	2024-10	2X6
B0020	CALIBRATORE Toxocara IgG	978	2024-10	1
C0020	CONTROL SERUM Toxocara IgG	978	2024-10	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	555	≥ 100
Control (Index)	2,4	1,6-3,0

CALIBRATOR



9	3	3	9	7	8	6	0	0	0	0	0	0	4	4	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature HB

CONTROL SERUM

9	3	3	9	7	8	6	1	6	3	0	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature PB



This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents. *[Signature]*

DATE 08.11.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.