

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME CYTOMEGALOVIRUS IgG -CHORUS

Code 81010

Lot No. 979



2024-10



2023-10

REAGENTS

Label	Description	Lot.	Exp.	NO
D2161	Cytomegalovirus IgG	979	2024-10	6x6
B2162	CALIBRATORE Cytomegalovirus IgG	979	2024-10	1
C2161	CONTROL SERUM Cytomegalovirus IgG	979	2024-10	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	406	≥ 100
Control Serum (IU/mL)	9,4	5,0-15,0
Negative Control (IU/mL)	<0,50	<0,5

CALIBRATOR

0	0	1	9	7	9	6	6	2	5	3	6	2	8	5	6	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature *RB*




CONTROL SERUM

0	0	1	9	7	9	6	2	5	4	6	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature *RB*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *[Signature]*

DATE 07/11/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.