

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME dsDNA-M - CHORUS.

Code 86034

Lot No. 979

2025-01

2023-10

REAGENTS

Label	Description	Lot.	Exp.	NO
D7401	dsDNA-M	979	2025-01	6X6
B7401	CALIBRATORE dsDNA-M	979	2025-01	1
C7401	CONTROL SERUM dsDNA-M	979	2025-01	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	740	≥ 100
Control Serum (AU/mL)	50,5	39,6-87,5

CALIBRATOR

2	5	7	9	7	9	7	5	2	2	3	4	4	2	4	6	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature



CONTROL SERUM

2	5	7	9	7	9	7	1	7	3	1	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.

DATE 14/11/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.