

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

DIESSE

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME INFLUENZA B IgG -CHORUS

Code 81193

Lot No. 987



2024-11



2023-11

REAGENTS

Label	Description	Lot.	Exp.	NO
D7531	INFLUENZA B IgG	987	2024-11	6X6
B7531	CALIBRATORE INFLUENZA B IgG	987	2024-11	1
C7531	CONTROL SERUM INFLUENZA B IgG	987	2024-11	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	300	≥ 100
Control Serum (Ndx)	3,9	2,8-5,2
Negative Control (Ndx)	0,1	$\leq 0,7$


CALIBRATOR

0	7	2	9	8	7	6	0	0	0	0	0	2	5	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature *MB*



CONTROL SERUM

0	7	2	9	8	7	6	2	8	5	2	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature *MB*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *[Signature]*

DATE *30/11/2023*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.