



DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME DEAMIDATED GLIADIN-G - CHORUS

Code 86100

Lot No. 988

2024-11

2023-11

REAGENTS

Label	Description	Lot.	Exp.	NO
D8660	Deamidated Gliadin-G	988	2024-11	6X6
B8660	CALIBRATORE Deamidated Gliadin-G	988	2024-11	1
C8660	CONTROL SERUM Deamidated Gliadin-G	988	2024-11	1

CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	605	≥ 100
Control Serum (AU/mL)	49,3	31,2-93,8

CALIBRATOR

2	5	5	9	8	8	6	1	2	3	3	1	4	4	5	3	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *[Signature]*



CONTROL SERUM

2	5	5	9	8	8	6	1	6	3	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *[Signature]*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *[Signature]*

DATE 28-11-2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.