

QUALITY CONTROL CERTIFICATE

KIT NAME SYPHILIS SCREEN RECOMBINANT - CHORUS

Code 81100

Lot No. 971

2024-09

2023-09

REAGENTS

Label	Description	Lot.	Exp.	NO
D2541	Syphilis Screen Recombinant	971	2024-09	6X6
B2541	CALIBRATORE Syphilis Screen Recombinant	971	2024-09	1
C2541	CONTROL SERUM Syphilis Screen Recombinant	971	2024-09	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	450	≥ 100
Control Serum (INDEX)	2,1	1,6-3,0
Negative Control (INDEX)	0,2	$<0,7$

CALIBRATOR

0	3	2	9	7	1	6	0	0	0	0	0	0	4	4	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature

[Signature]



CONTROL SERUM

0	3	2	9	7	1	6	1	6	3	0	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature

[Signature]



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

[Signature]

DATE 29/11/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.