

DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

### QUALITY CONTROL CERTIFICATE

KIT NAME BORRELIA IgM - CHORUS

Code 81047

Lot No. 963



2024-08



2023-08

### REAGENTS

Label	Description	Lot.	Exp.	NO
D2310	BORRELIA IgM	963	2024-08	6X6
B2310	CALIBRATORE BORRELIA IgM	963	2024-09	1
C2310	CONTROL BORRELIA IgM	963	2024-09	1

### CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	360	$\geq 100$
Control (Index)	4,1	2,8-5,2

### CALIBRATOR

9	0	5	9	6	3	6	0	0	0	0	0	2	5	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



### CONTROL SERUM


9	0	5	9	6	3	6	2	8	5	2	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



*This batch has been manufactured and checked according to quality requirements.  
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 28/11/2023

### WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.