

QUALITY CONTROL CERTIFICATE

KIT NAME MEASLES IgG CHORUS

Code 81072

Lot No. 989

2025-02

2023-11

REAGENTS

| Label | Description | Lot. | Exp. | NO |
|-------|---------------------------|------|---------|-----|
| D2431 | MEASLES IgG | 989 | 2025-02 | 6X6 |
| B2432 | CALIBRATORE MEASLES IgG | 989 | 2025-02 | 1 |
| C2431 | CONTROL SERUM MEASLES IgG | 989 | 2025-02 | 1 |

CONTROLS

| | Found results | Expected results |
|--|---------------|------------------|
|--|---------------|------------------|

| | | |
|--------------------------|-----|------------|
| Calibrator (mOD) | 490 | ≥ 100 |
| Control Serum (Index) | 3,5 | 2,2-4,1 |
| Negative Control (Index) | 0,0 | $< 0,7$ |

CALIBRATOR

0 0 9 9 8 9 7 0 0 0 0 0 3 2 0 0 0 0

Label checked with barcode reader

YES ☒

Operator Signature *FS*



CONTROL SERUM

0 0 9 9 8 9 7 2 2 4 1 0 0 0 0 0 0 1

Label checked with barcode reader

YES ☒

Operator Signature *FS*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *FS*

DATE 01/12/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.