

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME TETANUS IgG - CHORUS

Code 81270

Lot No. 987

2024-11

2023-11

REAGENTS

Label	Description	Lot.	Exp.	NO
D9660	Tetanus IgG	987	2024-11	6X6
B9660	CALIBRATORE Tetanus IgG	987	2024-11	1
C9660	CONTROL SERUM Tetanus IgG	987	2024-11	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	574	≥ 100
Control (mIU/mL)	743,6	431,8-1000,0


CALIBRATOR

9 0 2 9 8 7 6 3 2 1 3 3 4 0 5 0 0 0

Label checked with barcode reader

YES ☒

Operator Signature




CONTROL SERUM

9 0 2 9 8 7 6 1 8 3 3 0 0 0 0 0 0 1

Label checked with barcode reader

YES ☒

Operator Signature



**This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.**

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 11/12/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.