

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME ENZYWELL EPSTEIN BARR EBNA IgG

Code	91057		
LOT.NO	104-A	2024-12	2023-10

REAGENTS

CODE	LABEL	DESCRIPTION	LOT	EXP	NO
PF93071	054800	MICROPLATE	120	2025-03	1
PF93509	005470	CONJUGATE	104	2025-01	1
PF93910	1036	NEGATIVE CONTROL	177	2025-03	1
PF92073	005390	POSITIVE CONTROL	096	2025-04	1
PF91873	005450	CUT OFF CONTROL	096	2025-04	1
PF93619	1034	SUBSTRATE-HS	380	2025-03	1
PF93621	006030	DILUENT 10	164	2024-12	1
PF93603	1035	WASHING BUFFER 10x	462	2025-02	1
PF93602	1032/RISK0	STOP SOLUTION	218-1	2025-06	1
PF91083	006840	SORBENT G	094	2024-12	1

CONTROLS

	FOUND RESULTS	EXPECTED RESULTS
NEGATIVE CONTROL	0.014	$\leq 0,6$ Cut Off
CUT OFF CONTROL	0.396	≥ 0.200
POSITIVE CONTROL	0.988	$\geq 1,5$ Cut Off
CV %	8%	$\leq 15\%$

*Evaluation of the method according to the package inset in the attached document
This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED BY: RESPONSIBLE HEAD, Quality Control Reagents

DATE: 25.6.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.

IO 09- 022-V



CODE 91057
CODE _____

QUALITY CONTROL CERTIFICATE

Annex: Method Evaluation

NAME KIT EPSTEIN BARR EBNA IgG
LOT. 104
DATE OF TEST 18/10/2023

METHOD: ACCORDING TO THE PACKAGE INSERT

1. SOLID PHASE	FOUND		EXPECTED
A) UNIFORMITY BETWEEN PLATES			
1st PLATE	OD 450 nm	1.941	
	CV%	7%	< 20%
2nd PLATE	OD 450 nm	N/A	
	CV%	N/A	< 20%
B) UNIFORMITY WITHIN PLATE			
	OD 450 nm	2.148	
	CV%	6%	< 15%
2. BULK CONTROL date: 12/10/2023			
A) Average OD of Cut Off Control	OD 450 nm	0.457	≥ 0.200
B) Index of Negative and Positive Sera		accepted	According to specification
C) Sensitivity with working calibrators		accepted	According to specification
3. FINAL CONTROL date: 18/10/2023			
A) Blank	OD 450 nm	0.039	≤ 0.100
B) Average OD of Negative Control	OD 450 nm	0.014	≤ 0,6 Cut Off
C) Average OD of Cut Off Control	OD 450 nm	0.396	≥ 0.200
D) Average OD of Positive Control	OD 450 nm	0.988	≥ 1,5 Cut Off
E) HOMOGENEITY OF CUT OFF			
	No.	6	
	OD 450 nm	0.434	
	CV%	8%	≤ 15%
F) PRECISION OF FILLING PROCESS			

☒ SATISFACTORY

☐ UNSATISFACTORY

FINAL RESULT: ACCEPTED

Date: 18.10.2023

Signed: 
Released: Q.C.Manager