

DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

## QUALITY CONTROL CERTIFICATE

KIT NAME snRNP-C - CHORUS

Code 86092

Lot No. 996

2024-12

2023-12

### REAGENTS

Label	Description	Lot.	Exp.	NO
D7441	snRNP-C	996	2024-12	6X6
B7441	CALIBRATORE snRNP-C	996	2024-12	1
C7441	CONTROL SERUM snRNP-C	996	2024-12	1

### CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	360	$\geq 100$
Control Serum (AU/mL)	67,9	33,0-85,5

### CALIBRATOR

1	9	1	9	9	6	6	2	1	8	2	8	3	8	3	9	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES ☒ Operator Signature MB



### CONTROL SERUM

1	9	1	9	9	6	6	1	9	3	5	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature MB



This batch has been manufactured and checked according to quality requirements.  
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents. MB

DATE 14/12/2023

### WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.