

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME a-TG - CHORUS

Code 86072

Lot No. 971



2024-09



2023-12

REAGENTS

Label	Description	Lot.	Exp.	NO
D3571	a-TG	971	2024-09	6X6
B3571	CALIBRATORE a-TG	971	2024-12	1
C3571	CONTROL SERUM a-TG	971	2024-12	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	321	≥ 100
Control Serum (IU/mL)	318,1	266,0-517,6

CALIBRATOR


3	0	0	9	7	1	6	1	3	0	6	4	3	3	7	7	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature MB



CONTROL SERUM


3	0	0	9	7	1	6	2	2	4	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature MB



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

[Signature]

DATE 18/12/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.