

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

DIESSE

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME VARICELLA IgM CHORUS

Code 81079

Lot No. 996

2025-03

2023-12

REAGENTS

Label	Description	Lot.	Exp.	NO
D2481	VARICELLA IgM			
B2481	CALIBRATORE VARICELLA IgM	996	2025-03	6x6
C2481	CONTROL SERUM VARICELLA IgM	996	2025-03	1
		996	2025-03	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	499	≥ 100
Control Serum (Index)	2,6	1,8-3,3
Negative Control (Index)	0,1	< 0,7

CALIBRATOR

1	3	4	9	9	6	7	0	0	0	0	0	4	0	0	0	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature MB




CONTROL SERUM

1	3	4	9	9	6	7	1	8	3	3	0	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature MB



This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 29/12/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.