

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

DIESSE

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME TETANUS IgG - CHORUS

Code 81270/12

Lot No. 988



2024-11



2023-11

REAGENTS

Label	Description	Lot.	Exp.	NO
D9660	Tetanus IgG	988	2024-11	2X6
B9660	CALIBRATORE Tetanus IgG	988	2024-11	1
C9660	CONTROL SERUM Tetanus IgG	988	2024-11	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	583	≥ 100
Control (mIU/mL)	664,2	431,8-1000,0

CALIBRATOR

9	0	2	9	8	8	6	3	2	1	3	3	4	0	5	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader

YES



Operator Signature

MB



CONTROL SERUM

9	0	2	9	8	8	6	1	8	3	3	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature

MB



This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 05/01/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.