

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME MYCOPLASMA IgA - CHORUS

Code 81033

Lot No. 996



2024-12



2023-12

REAGENTS

Label	Description	Lot.	Exp.	NO
D751	MYCOPLASMA IgA	996	2024-12	6X6
B751	CALIBRATORE MYCOPLASMA IgA	996	2024-12	1
C751	CONTROL SERUM MYCOPLASMA IgA	996	2024-12	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	413	≥ 100
Control Serum (AU/mL)	22,9	15,0-41,9
Negative Control (AU/mL)	<10,0	<10


CALIBRATOR

0	6	8	9	9	6	6	9	1	2	1	8	6	8	4	9	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature *SB*



CONTROL SERUM

0	6	8	9	9	6	6	1	0	1	9	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature *SB*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *SB*

DATE *11.01.2024*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.