

DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

## QUALITY CONTROL CERTIFICATE

KIT NAME TICK-BORNE ENCEPHALITIS VIRUS IgG - CHORUS

Code 81276

Lot No. 999

2024-12

2023-12

### REAGENTS

Label	Description	Lot.	Exp.	NO
D0140	TICK-BORNE ENCEPHALITIS VIRUS IgG	999	2024-12	6X6
B0141	CALIBRATORE TICK-BORNE ENCEPHALITIS VIRUS IgG	999	2024-12	1
C0141	CONTROL SERUM TICK-BORNE ENCEPHALITIS VIRUS IgG	999	2024-12	1

### CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	520	$\geq 100$
Control (AU/mL)	47,9	32,5-56,0


### CALIBRATOR

9	6	6	9	9	9	6	1	1	8	4	3	4	7	7	7	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES

Operator Signature



### CONTROL SERUM

9	6	6	9	9	9	6	1	5	2	8	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES

Operator Signature



This batch has been manufactured and checked according to quality requirements.  
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.