

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

DIESSE

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME CARDIOLIPIN-M - CHORUS

Code 86048

Lot No. 001



2025-01



2024-01

REAGENTS

Label	Description	Lot.	Exp.	NO
D3331	CARDIOLIPIN-M	001	2025-01	6X6
B3331	CALIBRATORE CARDIOLIPIN-M	001	2025-01	1
C3331	CONTROL SERUM CARDIOLIPIN-M	001	2025-01	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	710	≥ 100
Control Serum (MPLU/mL)	52,3	30,0-100,0

CALIBRATOR

3	3	6	0	0	1	6	5	1	9	2	8	3	8	3	5	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature

MB

CONTROL SERUM

3	3	6	0	0	1	6	1	9	3	5	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader

YES



Operator Signature

MB



This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

[Signature]

DATE 26.01.2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.