

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SIENA) – Italy – Strada dei Laghi, 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME: PARVOVIRUS B19 IgM - CHORUS

Code	81152		
Lot No.	996	2024-12	2024-01

REAGENTS

Label	Description	Lot.	Exp	No
D0250	Parvovirus B19 IgM	996	2024-12	6X6
B0250	CALIBRATORE Parvovirus B19 IgM	996	2024-12	1
C0250	CONTROL SERUM Parvovirus B19 IgM	996	2024-12	1

CONTROLS

	Found results	Expected resu
--	---------------	---------------

CALIBRATOR (MOD)	270	≥100
CONTROL SERUM (INDEX)	3,5	2,7-5,0

CALIBRATOR


0	2	2	9	9	6	6	0	0	0	0	0	2	6	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



CONTROL SERUM


0	2	2	9	9	6	6	2	7	5	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



The test has been carried out in accordance with the package insert

TESTED BY: Chiara Bruschelli

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 23.01.2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.