

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME CALPROTECTIN K - CHORUS

Code	86601		
Lot No.	001	2025-01	2024-01

REAGENTS

Label	Description	Lot.	Exp.	NO
D9970	CALPROTECTIN	001	2025-01	2X6
B9970	CALIBRATORE CALPROTECTIN	001	2025-01	1
C9970	CONTROL CALPROTECTIN	001	2025-01	1
G9502	SAMPLE DILUENT	001	2026-01	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	360	≥ 100
Control Serum (mg/Kg)	630,2	396,0-1000,0

CALIBRATOR

9	2	9	0	0	1	6	3	2	6	5	1	2	2	5	9	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature FB



CONTROL SERUM

9	2	9	0	0	1	6	3	2	5	9	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature FB



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents FB

DATE 26.01.2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.