

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME **HELICOBACTER PYLORI IgG -CHORUS**

Code **81060**

Lot No. **002**



2025-01



2024-01

REAGENTS

Label	Description	Lot.	Exp.	NO
D2391	Helicobacter Pylori IgG	002	2025-01	6X6
B2392	CALIBRATORE Helicobacter Pylori IgG	002	2025-01	1
C2391	CONTROL SERUM Helicobacter Pylori IgG	002	2025-01	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	381	≥ 100
Control Serum (UA/mL)	101,7	75,0-188,9
Negative Control (UA/mL)	<5	<5

CALIBRATOR

0	0	8	0	0	2	6	1	2	0	3	2	2	7	5	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature

FB



CONTROL SERUM

0	0	8	0	0	2	6	2	6	4	8	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature

FB



This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

Bruno

DATE

26.01.2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.