Pag 1/1

DIESSE DIAGNOSTICA SENESE SpA

DIESSE

Sede Legale 20144 Milano (Italy) Via A. Solari 19 sc. 6

53035 Monteriggioni (SI), Italy, Strada dei Laghi 39 Tel. +39 0577319554 FAX +39 0577319020

| PRODUCT | NAME: | VACU-TEC |
|---------|-------|----------|
|         |       |          |

**Code** 10208

**Lot No.** 4030

2024-11

M 2024-01

Reference IO-09-88

| 1101  | crence 10-00-00  |                 |     |
|-------|--|-----------------|-----|
| ON    | LINE QUALITY CONTROLS (model CDP not enclosed)                                       | QC Passed       |     |
|       | sodium citrate solution concentration  | <b>₩</b>        |     |
| •     | sodium citrate volume  | Γ               |     |
|       | vacuum level (*)   | Ø               |     |
|       | correct placement of the cap   | I♥              |     |
| •     | no extra body inside the product   | <b>₽</b>        |     |
| (*) ( | Only for VACU-CODE and VACU-TEC tubes  | च               |     |
| PAC   | CKAGING QUALITY CONTROL (model CDP not enclosed)                                     | QC Passed       |     |
| *     | control of the number of the product for each pack                                   | <b>₩</b>        |     |
| QU    | ALITY CONTROLS ON FINISHED PRODUCT (model CDP not enclosed                           | d)<br>QC Passed |     |
| •     | pack integrity   | Ι <del>Σ</del>  |     |
| •     | data printed on the product label  | ट्य             |     |
|       |  | QC Passed       | N/A |
| Irra  | rradiation treatment Certificate n° date   |                 | 17  |
|       |  | QC Passed       | N/A |
|       | nma rays treatment process in compliance with the specifications ne supply Agreement | e               | (e  |
| Doc   | ument number: ODV N.A.   |                 |     |

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED

RESPONSIBLE HEAD, Quality Control Reagents.

DATE DE DE BLY