

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME **HELICOBACTER PYLORI IgG -CHORUS**

Code **81060**
Lot No. **003** **2025-01** **2024-01**

REAGENTS

Label	Description	Lot.	Exp.	NO
D2391	Helicobacter Pylori IgG	003	2025-01	6X6
B2392	CALIBRATORE Helicobacter Pylori IgG	003	2025-01	1
C2391	CONTROL SERUM Helicobacter Pylori IgG	003	2025-01	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	398	≥ 100
Control Serum (UA/mL)	145,3	75,0-188,9
Negative Control (UA/mL)	<5	<5

CALIBRATOR

0 0 8 0 0 3 6 1 2 0 3 2 2 7 5 0 0 0

Label checked with barcode reader YES ☒ Operator Signature *HB*

CONTROL SERUM

0 0 8 0 0 3 6 2 6 4 8 0 0 0 0 0 0 1

Label checked with barcode reader YES ☒ Operator Signature *HB*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents. *[Signature]*

DATE *07-02-2024*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.