

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

DIESSE

Sede Legale
20144 Milano (Italy)

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME 88000-C204 - AMOXICILLIN-HSA

Code	88000-C204		
Lot No.	001	2025-01	2024-02

REAGENTS

Label	Description	Lot.	Exp	No
D3741	88000- C204 AMOXICILLIN-HSA	001	2025-01	2X6
B3740	CALIBRATORE 88000- C204 AMOXICILLIN-HSA	001	2025-02	1
C3740	CONTROL SERUM 88000-C204 AMOXICILLIN-HSA	001	2025-02	1

CONTROLS

Found results	Expected results
---------------	------------------

Calibrator (mOD)	284	> 100
Control Serum (KU/L)	8,2	5,8-14,3


CALIBRATOR

6	9	7	0	0	1	6	5	2	1	6	0	2	6	8	1	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature




CONTROL SERUM

6	9	7	0	0	1	6	2	7	5	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution as Research Use Only (RUO) product.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 08/02/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.