

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME CHORUS PHOSPHOLIPID SCREEN IgM

Code 86108

Lot No. 001



2025-01



2024-01

REAGENTS

Label	Description	Lot.	Exp.	NO
D4020	Phospholipid Screen IgM	001	2025-01	6X6
B4020	CALIBRATORE Phospholipid Screen IgM	001	2025-01	1
C4020	CONTROL SERUM Phospholipid Screen IgM	001	2025-01	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	680	≥ 100
Control (AU/ml)	39,1	22,5-80,0

CALIBRATOR


9	7	5	0	0	1	6	1	1	9	2	7	3	6	3	6	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *FB*



CONTROL SERUM

9	7	5	0	0	1	6	2	0	3	6	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *FB*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 09/02/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.