

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

DIESSE

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME CYTOMEGALOVIRUS IgM-CHORUS

Code 81011

Lot No. 998

2025-03

2023-12

REAGENTS

Label	Description	Lot.	Exp.	NO
D2171	Cytomegalovirus IgM	998	2025-03	6X6
B2171	CALIBRATORE Cytomegalovirus IgM	998	2025-03	1
C2171	CONTROL SERUM Cytomegalovirus IgM	998	2025-03	1
G2171	CONFIRMATORY TEST	998	2025-03	5X1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	497	≥ 100
Control Serum (INDEX)	1,8	1,4-2,6
Negative Control (INDEX)	0,1	$< 0,7$

CALIBRATOR


0 7 9 9 9 8 7 0 0 0 0 0 5 1 0 0 0 0

Label checked with barcode reader

YES



Operator Signature *FB*



CONTROL SERUM


0 7 9 9 9 8 7 1 4 2 6 0 0 0 0 0 0 1

Label checked with barcode reader

YES



Operator Signature *FB*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *[Signature]*

DATE 13/02/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.