

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

DIESSE

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME LEGIONELLA PNEUMOPHILA IgM -CHORUS

Code	81093
Lot No.	007
	2024-11
	2024-02

REAGENTS

Label	Description	Lot.	Exp.	NO
D7391	LEGIONELLA PNEUMOPHILA IgM	007	2024-11	6X6
B7391	CALIBRATORE L. PNEUMOPHILA IgM	007	2024-11	1
C7391	CONTROL SERUM L. PNEUMOPHILA IgM	007	2024-11	1

CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	312	≥ 100
Control Serum (Index)	3,4	2,5-4,7
Negative Control (Index)	0,1	$<0,7$

CALIBRATOR

1	3	6	0	0	7	5	0	0	0	0	0	2	8	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *SS*



CONTROL SERUM

1	3	6	0	0	7	5	2	5	4	7	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *SS*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents... *SS*

DATE *14.02.2024*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.