

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME CHORUS Aspergillus Galactomannan Ag

Code 81178

Lot No. 007



2025-02



2024-02

REAGENTS

Label	Description	Lot.	Exp.	NO
D4070	ASPERGILLUS Ag	007	2025-02	6X6
B4070	CALIBRATORE ASPERGILLUS Ag	007	2025-02	1
C4070	CONTROL SERUM ASPERGILLUS Ag	007	2025-02	2
G4070	SAMPLE DILUENT ASPERGILLUS Ag	LP2312	2025-06	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	687	≥ 100
Control Serum (Index)	1,9	1,3-2,5

CALIBRATOR

9	7	7	0	0	7	6	0	0	0	0	0	5	3	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *FB*



CONTROL SERUM

9	7	7	0	0	7	6	1	3	2	5	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *FB*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *Antonello*

DATE *27/02/2024*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.