

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME CHORUS PHOSPHOLIPID SCREEN IgG

Code 86106

Lot No. 007 2025-02 2024-02

REAGENTS

Label	Description	Lot.	Exp.	NO
D4010	Phospholipid Screen IgG	007	2025-02	6X6
B4010	CALIBRATORE Phospholipid Screen IgG	007	2025-02	1
C4010	CONTROL SERUM Phospholipid Screen IgG	007	2025-02	1

CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	570	≥ 100
Control (AU/ml)	29,9	18,0-45,7

CALIBRATOR

9	7	4	0	0	7	6	5	1	5	2	4	4	9	4	5	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES ☒ Operator Signature HB



CONTROL SERUM

9	7	4	0	0	7	6	1	4	2	7	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature YB



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents Agus

DATE 07/03/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.