

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME HERPES SIMPLEX 1+2 IgM -CHORUS

Code	81021
Lot No.	009
	2025-02
	2024-02

REAGENTS

Label	Description	Lot.	Exp.	NO
D2221	Herpes simplex 1+2 IgM	009	2025-02	6x6
B2221	CALIBRATORE Herpes simplex 1+2 IgM	009	2025-02	1
C2221	CONTROL SERUM Herpes simplex 1+2 IgM	009	2025-02	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	496	≥ 100
Control Serum (INDEX)	2,0	1,3-2,5
Negative Control (INDEX)	0,0	$<0,7$

CALIBRATOR

1	5	6	0	0	9	6	0	0	0	0	0	5	4	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *FB*



CONTROL SERUM

1	5	6	0	0	9	6	1	3	2	5	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *FB*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *Blue*

DATE 12/03/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.