

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

DIESSE

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME TOXOPLASMA IgM - CHORUS

Code	81041
Lot No.	998
	2024-12
	2023-12

REAGENTS

Label	Description	Lot.	Exp.	NO
D2281	TOXOPLASMA IgM	998	2024-12	6X6
B2281	CALIBRATORE TOXOPLASMA IgM	998	2024-12	1
C2281	CONTROL SERUM TOXOPLASMA IgM	998	2024-12	1


CONTROLS

	Found results	Expected results
Calibrator (mOD)	524	≥ 100
Control Serum (INDEX)	1,3	0,9-1,7
Negative Control (INDEX)	0,0	$< 0,7$

CALIBRATOR

1	3	1	9	9	8	6	0	0	0	0	0	7	6	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES ☒ Operator Signature *FB*



CONTROL SERUM

1	3	1	9	9	8	6	0	9	1	7	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *FB*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *[Signature]*

DATE 13/03/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.