

DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

### QUALITY CONTROL CERTIFICATE

KIT NAME EPSTEIN-BARR EARLY ANTIGEN IgM

Code 81059

Lot No. 007



2025-02



2024-02

### REAGENTS

Label	Description	Lot.	Exp.	NO
D2382	Epstein-Barr Early Antigen IgM	007	2025-02	6x6
B2382	CALIBRATORE Epstein-Barr Early Antigen IgM	007	2025-02	1
C2382	CONTROL SERUM Epstein-Barr Early Antigen IgM	007	2025-02	1

### CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	340	$\geq 100$
Control Serum (Index)	2,8	1,9-3,5
Negative Control (Index)	0,0	$< 0,7$

### CALIBRATOR

0	2	9	0	0	7	6	0	0	0	0	0	3	7	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader

YES



Operator Signature

FB



### CONTROL SERUM

0	2	9	0	0	7	6	1	9	3	5	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature

FB



*This batch has been manufactured and checked according to quality requirements.  
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

FB

DATE 20/03/2024

### WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.