

DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020

**DIESSE**

Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME MEASLES IgG CHORUS

Code 81072

Lot No. 990



2025-02



2023-11

REAGENTS

Label	Description	Lot.	Exp.	NO
D2431	MEASLES IgG	990	2025-02	6X6
B2432	CALIBRATORE MEASLES IgG	990	2025-02	1
C2431	CONTROL SERUM MEASLES IgG	990	2025-02	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	410	$\geq 100$
Control Serum (Index)	3.5	2.4-4.4
Negative Control (Index)	0.0	$< 0.7$

CALIBRATOR


0 0 9 9 9 0 7 0 0 0 0 0 3 0 0 0 0 0 0

Label checked with barcode reader

YES



Operator Signature *Green*



CONTROL SERUM


0 0 9 9 9 0 7 2 4 4 4 0 0 0 0 0 0 0 1

Label checked with barcode reader

YES



Operator Signature *Green*



*This batch has been manufactured and checked according to quality requirements.  
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *Green*

DATE 21.03.2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.