

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME PARVOVIRUS B19 IgM-CHORUS

Code 81152

Lot No. 017



2025-03



2024-03

REAGENTS

Label	Description	Lot.	Exp	No
D0250	Parvovirus B19 IgM	017	2025-03	6X6
B0250	CALIBRATORE Parvovirus B19 IgM	017	2025-03	1
C0250	CONTROL SERUM Parvovirus B19 IgM	017	2025-03	1

CONTROLS

Found results	Expected results
---------------	------------------

CALIBRATOR (mOD)	504	≥100
CONTROL SERUM (INDEX)	2,6	1,8-3,4

CALIBRATOR

0	2	2	0	1	7	6	0	0	0	0	0	3	9	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *FS*



CONTROL SERUM

0	2	2	0	1	7	6	1	8	3	4	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *FS*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE *28/03/2024*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.