

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

DIESSE

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME ANTI-CCP - CHORUS

Code 86094

Lot No. 018



2025-03



2024-03

REAGENTS

Label	Description	Lot.	Exp.	NO
D8671	ANTI-CCP	018	2025-03	6x6
B8671	CALIBRATORE ANTI-CCP	018	2025-03	1
C8671	CONTROL SERUM ANTI-CCP	018	2025-03	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	543	≥ 100
Control Serum (AU/mL)	49,7	36,5-76,3

CALIBRATOR

2 5 6 0 1 8 6 2 2 3 4 2 4 0 6 3 0 0

Label checked with barcode reader

YES



Operator Signature

FB



CONTROL SERUM

2 5 6 0 1 8 6 1 8 3 3 0 0 0 0 0 0 1

Label checked with barcode reader

YES



Operator Signature

FB



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

[Signature]

DATE 07/04/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.