

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME INFLUENZA B IgA -CHORUS

Code 81195

Lot No. 001



2025-01



2024-03

REAGENTS

Label	Description	Lot.	Exp.	NO
D7541	INFLUENZA B IgA	001	2025-01	6X6
B7541	CALIBRATORE INFLUENZA B IgA	001	2025-01	1
C7541	CONTROL SERUM INFLUENZA B IgA	001	2025-01	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	630	≥ 100
Control Serum (Ndx)	1,3	1,0-1,9
Negative Control (Ndx)	0,4	$\leq 0,7$

CALIBRATOR


0	7	1	0	0	1	6	0	0	0	0	0	6	8	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature 



CONTROL SERUM


0	7	1	0	0	1	6	1	0	1	9	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader


YES



Operator Signature 



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 02-04-2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive