

DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020

**DIESSE**

Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME ANA-Screen - CHORUS

Code 86014

Lot No. 017



2025-03



2024-03

REAGENTS

Label	Description	Lot.	Exp.	NO
D3141	ANA-Screen	017	2025-03	6X6
B3141	CALIBRATORE ANA-Screen	017	2025-03	1
C3141	CONTROL SERUM ANA-Screen	017	2025-03	1

CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	350	$\geq 100$
Control Serum (Ndx)	3,2	2,1-4,0

CALIBRATOR

1	1	5	0	1	7	6	0	0	0	0	0	3	3	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature FB



CONTROL SERUM

1	1	5	0	1	7	6	2	1	4	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature FB



*This batch has been manufactured and checked according to quality requirements.  
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents head

DATE 03/04/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.