

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME CLOSTRIDIUM DIFFICILE GDH - CHORUS

Code	81168
Lot No.	018
	2025-03
	2024-03

REAGENTS

Label	Description	Lot.	Exp.	NO
D1880	CLOSTRIDIUM DIFFICILE GDH	018	2025-03	2X6
B1880	CALIBRATORE C.DIFFICILE GDH	018	2025-03	1
C1880	CONTROL SERUM C.DIFFICILE GDH	018	2025-03	1
G1880	C.DIFFICILE Stool Diluent/Negative Control	E0199500	2025-07	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	290	≥ 100
Control + (Index)	3,9	2,8-5,2
Stool Diluent/Negative Control (Index)	0,3	$<0,5$

CALIBRATOR

9	3	5	0	1	8	6	0	0	0	0	0	2	5	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *FB*



CONTROL SERUM

9	3	5	0	1	8	6	2	8	5	2	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *FB*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *looney*

DATE 03/04/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.