

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME Vitamin D Total- CHORUS

Code 86900

Lot No. 009



2025-02



2024-03

REAGENTS

Label	Description	Lot.	Exp.	NO
D9702	25OH Vitamin D Total	009	2025-02	6x6
B9703	CALIBRATOR 25OH Vitamin D Total	009	2026-02	1
C9703	CONTROL SERUM 25OH Vitamin D Total	009	2026-02	4

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	512	≥ 100
CONTROL SERUM (ng/ml)	34,5	32,2-39,2

CALIBRATOR


8 3 5 0 0 9 6 3 3 3 3 7 6 7 5 6 0 0

Label checked with barcode reader

YES



Operator Signature *FB*



CONTROL SERUM


8 3 5 0 0 9 6 1 1 2 0 0 0 0 0 0 0 1

Label checked with barcode reader

YES



Operator Signature *FB*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *FB*

DATE 19/04/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.