

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME BETA 2-GLYCOPROTEIN-G - CHORUS

Code 86050

Lot No. 016

2025-03

2024-03

REAGENTS

Label	Description	Lot.	Exp.	NO
D3351	BETA 2-GLYCOPROTEIN-G	016	2025-03	6X6
B3351	CALIBRATORE BETA 2-GLYCOPROTEIN-G	016	2025-03	1
C3351	CONTROL SERUM BETA 2-GLYCOPROTEIN-G	016	2025-03	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	548	≥ 100
Control Serum (AU/mL)	34,6	26,1-54,2

CALIBRATOR

2	0	6	0	1	6	6	2	2	9	4	1	4	4	5	8	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *[Signature]*



CONTROL SERUM

2	0	6	0	1	6	6	1	6	3	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *[Signature]*



This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *[Signature]*

DATE 15.04.2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.