

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

DIESSE

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME CYTOMEGALOVIRUS IgM-CHORUS

Code 81011

Lot No. 024



2025-07



2024-04

REAGENTS

Label	Description	Lot.	Exp.	NO
D2171	Cytomegalovirus IgM	024	2025-07	6X6
B2171	CALIBRATORE Cytomegalovirus IgM	024	2025-07	1
C2171	CONTROL SERUM Cytomegalovirus IgM	024	2025-07	1
G2171	CONFIRMATORY TEST	024	2025-07	5X1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	485	≥ 100
Control Serum (INDEX)	2,2	1,5-2,9
Negative Control (INDEX)	0,0	$< 0,7$


CALIBRATOR

0	7	9	0	2	4	7	0	0	0	0	0	4	6	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature *FB*




CONTROL SERUM

0	7	9	0	2	4	7	1	5	2	9	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature *FB*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *Abene*

DATE 29/04/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.